

CERTIFICATE OF ANALYSIS



PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2021-05
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0596041	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.57	--	0.37	PASS
Intercept	---	0.12	--	0.07	PASS
r	0.990	---	--	0.996	PASS
LOW RANGE					
Slope	2.8	6.9	--	4.4	PASS
Intercept	---	0.21	--	0.11	PASS
r	0.990	---	--	0.999	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0595845	3 mL
COAMATIC FVIII S-2765 + I-2581	N0596036	6 mL
COAMATIC FVIII BUFFER	N0596035	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Nehal Patel *N.M. Patel*
 PRINT SIGNATURE

DATE: 6/14/2019

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*
 PRINT SIGNATURE

DATE: 6/14/19
 6/14/2019

TITLE: Technical Supervisor