



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS


DQ7C00AZ

TECHNOTHROMBIN<sup>®</sup> TGA RB 5x0.5mL

REF 5006209

LOT 7C91B00.02

 2021-01-31

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RB	7C91B00	2021-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z92B00	150 - 250 200.8 OK

Durchführung von  
Prepared by

Michael KAFKA 

21. OKT. 2019

Datum / Date:

Freigabe von  
Approved by

  
Eva Maria HAGER

21. OKT. 2019

Datum / Date: