



| | | |
|---|---|----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ8F00AZ |
|---|---|----------|

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230

LOT 8F91B00.03

 2021-04-30

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|--------------------|---------|---|
| 50x1.5mL | TGA Substrat | 8F91B00 | 2021-04-30 |

| Analytische Prozedur Analytical Procedure | | | Anforderung Requirement | Ergebnis Result | |
|---|---|----------------------------|----------------------------|--------------------|----|
| Kontrollwiederfindung Recovery of controls | TGA Control High (nM Peak Thrombin) LOT 1Z92B00 | TGA RC High LOT 7F92B00 | 474 - 790 | 611.3 | OK |
| | TGA Control Low (nM Peak Thrombin) LOT 2894B00 | TGA RC High LOT 7F92B00 | 244 - 406 | 308.2 | OK |

Durchführung von
Prepared by

Michael KAFKA 

09. OKT. 2019

Datum / Date:

Freigabe von
Approved by

Riham EL KAMMAR 

10. OKT. 2019

Datum / Date: