
	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQ1Z00AZ</p>
---	---	---

TECHNOTHROMBIN[®] TGA Control High 5x1mL

REF 5006320

LOT 1Z92B00.01

 2020-08-31

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z92B00	2020-08-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC Low (nM Peak Thrombin) LOT 7D91B00	266 - 433	331.3 OK
	TGA Reagenz RC High (nM Peak Thrombin) LOT 7F92B00	474 - 790	611.3 OK

Durchführung von
Prepared by

Michael KAFKA

09. OKT. 2019

Datum / Date:

Freigabe von
Approved by

Riham EL KAMMAR

10. OKT. 2019

Datum / Date: