

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0700AZ
---	---	----------

## Ceveron<sup>®</sup> TGA CAL Set

REF      5006347  
LOT      0792B00.01  
      2020-11-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x3mL	Ceveron TGA CAL BUF	9S91B00	2021-02-28
1x0.5mL	Ceveron TGA CAL	5Y81B	2020-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration Gegen / against WHO 01/580		report	785	OK
Verwendbarkeit / usability Gerät / Analyzer: Ceveron: 102	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	1.0	OK
Kontrollwiederfindung Recovery of controls <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Y83B00	Peak Thrombin nM Ceveron <sup>®</sup> TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7R92B	206 – 344	284.4	OK
Kontrollwiederfindung Recovery of controls <span style="border: 1px solid black; padding: 2px;">LOT</span> 2784B00	Peak Thrombin nM Ceveron <sup>®</sup> TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7R92B	140 - 234	157.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

Michael KAFKA 

26. APR. 2019

Datum / Date:

**Freigabe von**  
**Approved by**

Fikret NASUFI 

26. APR. 2019

Datum / Date: