
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9P00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA BUF 5x3 mL

REF      5006360  
LOT      9P91B00.03  
      2021-05-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5 x 3 mL	TGA Buffer	9P91B00	2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / usability  Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve	$\geq 0.99$	0.99	OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Z91B00	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F92B00	417 – 695	617.0 OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Z91B00	TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D91B00	249 - 415	350.5 OK

Durchführung von  
Prepared by

Fikret NASUFI 

26. AUG. 2019

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA 

26. AUG. 2019

Datum / Date: