

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0W00.01
---	---	-----------


**Technochrom AT III Kit ~ 100 T.**

REF      5340224  
LOT      0W82B00.01  
      2020-12-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x	AT III Substrate Th-1	8B81B00	2021-03-31
1x	Technochrom ATIII Reagent A2	7S71B00	2020-12-31
2x25mL	Natrium Chlorid Lösung 0.9%	9581B00	2023-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha Chromogenic Substrate F IIa	CCN <span style="border: 1px solid black; padding: 2px;">LOT</span> 1P71B00	75.8 – 126.3 %	102.5	OK
	CCA <span style="border: 1px solid black; padding: 2px;">LOT</span> 3P71B00	33.3 – 55.4 %	43.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

Suzana GRESAK PODAKAR  


11. JUNI 2018

Datum / Date:

**Freigabe von**  
**Approved by**

  
**Christoph ENNE**

11. JUNI 2018

Datum / Date: