



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRA00AZ

TECHNOZYM[®] vWF:Ag ELISA Kit 96T

REF 5450201

LOT RA92B00.01

⌚ 2021-05-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:Ag Plate	RNF82B1	2021-06-30
1 x 0.3mL	Ag Conjugate	RNH81B1	2021-06-30
1x 80mL	Wash buffer Concentrate	RNA92B0	2021-06-30
1 x 90mL	Incubation buffer	RNB92B0	2021-07-31
1 x 12mL	TMB Substrate	RKN92B0	202-07-31
1 x 12mL	Stop solution	RKK92B0	2021-06-30
1 x 0.5mL	vWF Calibrator 1	1ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 2	2ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 3	3ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 4	4ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 5	5ND92B0	2021-06-30
1 x 0.5mL	vWF High Control	AND91B0	2021-05-31
1 x 0.5mL	vWF Low Control	BND91B0	2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.27	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND91B0)	0.92 – 1.54 IU/mL	1.07	OK
	Low Control (BND91B0)	0.22 – 0.50 IU/mL	0.36	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

19. SEP. 2019

Datum / Date:

Freigabe von
Approved by

Stefan STEINER *Stefan Steiner*

19. SEP. 2019

Datum / Date: