



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRB00AZ

TECHNOZYM[®] vWF:CBA ELISA Kit 96T

REF 5450301

LOT RB98B00.02

⌚ 2021-01-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	vWF:CBA Plate Collagen Typ III	RNC93B0	2021-06-30
1 x 0.3mL	CBA Conjugate	RNE91B0	2021-03-31
1x 80mL	Wash buffer Concentrate	RNA92B0	2021-06-30
1 x 90mL	Incubation buffer	RNB91B0	2021-03-31
1 x 12mL	TMB Substrate	RKN91B0	2021-01-31
1 x 12mL	Stop solution	RKK91B0	2021-02-28
1 x 0.5mL	vWF Calibrator 1	1ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 2	2ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 3	3ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 4	4ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 5	5ND92B0	2021-06-30
1 x 0.5mL	vWF High Control	AND91B0	2021-05-31
1 x 0.5mL	vWF Low Control	BND91B0	2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.84	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND91B0)	1.14 – 1.90 IU/mL	1.58	OK
	Low Control (BND91B0)	0.25 – 0.57 IU/mL	0.46	OK

Durchführung von
Prepared by

Michael KAFKA

16. OKT. 2019

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

17. OKT. 2019

Datum / Date: