



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRR00AZ

TECHNOZYM[®] ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601

LOT RR93B00.01

⌚ 2021-04-30

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	Adamts13 Ag Plate	RRS91B0	2021-04-30
1 x 0.3mL	Adamts13 Ag Conjugate	RQV91B0	2021-05-31
1x 80mL	Wash buffer Concentrate	RNA92B0	2021-06-30
1 x 90mL	Incubation buffer	RNB92B0	2021-07-31
1 x 12mL	TMB Substrate	RKN92B0	2021-07-31
1 x 12mL	Stop solution	RKK92B0	2021-06-30
1 x 0.5mL	Adamts13 Ag Calibrator 1	1RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 2	2RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 3	3RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 4	4RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 5	5RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag High Control	RRY91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Low Control	RRZ91B0	2021-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Δ Cal 1 – Cal 5	≥ 0.7	0.94	OK
Verwendbarkeit Usability	High Control (RRY91B0)	0.47 – 0.79 IU/mL	0.65	OK
	Low Control (RRZ91B0)	0.12 – 0.28 IU/mL	0.19	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

19. SEP. 2019

Datum / Date:

Freigabe von
Approved by

Eva Maria Hager
Eva Maria HAGER

19. SEP. 2019

Datum / Date: