



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

RU9DB00.02

TECHNOZYM® ADAMTS-13 Activity ELISA 96T.

REF	5450701-RUO
LOT	RU9DB00.02
	2021-01-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
2x	6mL	ADAMTS-13 Activity vWF73 Substrate	930	2021-01-31
1x	30mL	ADAMTS-13 Activity Reaction Buffer	930	2021-01-31
1x	12mL	ADAMTS-13 Activity HRP Conj Antibody	930	2021-01-31
1x	12mL	ADAMTS-13 Activity Color Reagent	930	2021-01-31
1x	53mL	ADAMTS-13 Activity Wash Buffer Conc	930	2021-01-31
1x	12mL	ADAMTS-13 Activity Stop Solution	930	2021-01-31
1x	n.a.	ADAMTS-13 Activity Plate 96T.	930	2021-01-31
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 1	1UT92B0	2021-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 2	2UT92B0	2021-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 3	3UT92B0	2021-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 4	4UT92B0	2021-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 6	6UT92B0	2021-06-30
1x	0.5mL	UY ADAMTS-13 Activity High Control	RUY92B0	2021-06-30
1x	0.5mL	UZ ADAMTS-13 Activity Low Control	RUZ92B0	2021-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 5	5UT92B0	2021-06-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Content	according to packaging SOP
	Vial label	confirmed
	Vial barcode	OK
	Box label	confirmed
	Box barcode	OK
	Package insert	confirmed
	Batch table	OK
Verwendbarkeit Usability	deltaOD Cal1-Cal6 ≥ 0.7	1.8 OK
Kontrollwiederfindung Recovery of controls	ADAMTS-13 Activity High Control LOT RUY92B0	0.49 - 0.81 IU/mL 0.66 IU/mL OK
	ADAMTS-13 Activity Low Control LOT RUZ92B0	0.14 - 0.32 IU/mL 0.21 IU/mL OK



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

RU9DB00.02

Durchgeföhrt von
Prepared by

Michael KAFKA

05. SEP. 2019

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

05. SEP. 2019

Datum / Date: