

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	FXa, 10 x 71 nkat	EXPIRATION DATE:	2021-08
PRODUCT NUMBER:	82098539		
LOT NUMBER:	N0596233		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	66	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	0.999	PASS
CONTROL RECOVERY	66.0	73.0	nkat/vial	71.4	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 6/24/2019 *6-24-19*
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel* DATE: 6/24/2019 *6-24-19*
 PRINT SIGNATURE

TITLE: Senior QC Analyst