



CERTIFICATE OF ANALYSIS

Human L-FABP ELISA kit

Catalog number: DPGLFABP

Lot number: -

Expiry date: Month Year

Size: 1 x 96 determinations

Method of Analysis:

According to manual HK404, edition .

Range of standards obtained by following the protocol:

25000 – 10000 – 4000 – 1600 – 640 – 256 – 102 – 0 pg/ml

Amount of Human L-FABP in standard vial: pg

Amount of Dilution buffer to reconstitute the standard vial: ml

Concentration Human L-FABP Standard after reconstitution: 25000 pg/ml

Quality Control Specifications and Results

Calibrator	Result	Specification
Concentration 25000 pg/ml	OD 450 nm:	OD 450 nm: > 1.7
Blank value	OD 450 nm: 0.	OD 450 nm: < 0.2
Detection level	Conform	Calculated mean of the blanks plus three times the standard deviation < 102 pg/ml

This product is for research use only.

We hereby certify that the above information is correct and that this product has been tested successfully by the Quality Control Department. This product was released for sale according to the existing specifications. This document has been produced electronically and is valid without a signature.

If you have any questions or comments regarding this product, please do not hesitate to contact us.

Manufactured for Diapharma.

Diapharma Group, Inc.
8948 Beckett Road
West Chester, OH 45069
USA

Customer Support: 1-800-526-5224
Technical Support: 1-800-447-3846
Fax: 513-860-9635
Email: info@diapharma.com