

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2302, 25 mg Substrate	PRODUCT #: 00082034039
EXP. DATE: 2022-07	LOT #: N0394771
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N0394771	2022-07	3 mL

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 4/11/2019 *4-11-19*

APPROVAL SIGNATURE: D. Tencza *D. Tencza* DATE: 4/11/2019 *4/11/19*