
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ6BM0AZ
---	---	----------

**TECHNOZYM® PAP Calibrator Set 5x0.5mL**

**REF**      TC12062  
**LOT**      6BM91B0.01  
      2021-01-31

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 0.5mL	Calibrator 1 PAP	1BM61B2	2021-01-31
1 x 0.5mL	Calibrator 2 PAP	2BM61B2	2021-01-31
1 x 0.5mL	Calibrator 3 PAP	3BM61B2	2021-01-31
1 x 0.5mL	Calibrator 4 PAP	4BM61B2	2021-01-31
1 x 0.5mL	Calibrator 5 PAP	5BM61B2	2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	1 – 2.5	1.55	OK
	OD Cal 5	< 0.2	0.05	OK
	High Control (ABM61B2)	98.9 – 164.9 ng/mL	129.74	OK
	Low Control (BBM61B2)	13.0 – 30.2 ng/mL	21.35	OK

**Durchführung von  
Prepared by**

**Fikret NASUFI** *Fikret Nasufi*

**13. SEP. 2019**

Datum / Date:

**Freigabe von  
Approved by**

**Riham EL KAMMAR**  
*El Kammar Riham*

**13. SEP. 2019**

Datum / Date: