

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7BM0AZ
---	---	----------

TECHNOZYM® PAP ELISA Control Set 2x0.5mL

REF TC12064
LOT 7BM91B0.01
 2021-01-31

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 0.5mL	PAP High Control	ABM61B2	2021-01-31
1 x 0.5mL	PAP Low Control	BBM61B2	2021-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / Usability	High Control (ABM61B2)	98.9 – 164.9 ng/mL	125.79 OK
	Low Control (BBM61B2)	13.0 – 30.2 ng/mL	20.43 OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

13. SEP. 2019

Datum / Date:

Freigabe von
Approved by

Riham EL KAMMAR
El Kammar Riham

13. SEP. 2019

Datum / Date: