



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0J00.01

TECHNOCHROM® C1-Inhibitor 30 T.

REF 5345003

LOT 0J93B00.01

🕒 2020-10-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x1mL	Coagulation Control N for C1-INH	1P81BC1H	2020-10-31
1x1mL	Coagulation Control A for C1-INH	3P82BC1H	2020-11-30
1x0.953 mL	Coagulation Reference for C1-INH	1R81BC1H	2020-10-31
1x3 mL	C1 Esterase	7B91B00	2021-03-31
1x3 mL	Substrate f. C1-Esterase	8C91B00	2023-09-30
1x25 mL	C1-INH Probenpuffer A	9M81B00	2023-10-31
1x20 mL	C1-INH Reaktionspuffer B	9N81B00	2023-11-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P81BC1H	85.1 – 141.8 % 117.5 OK
	CCA LOT 3P82BC1H	35.7 – 59.5 % 46.5 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK

Durchführung von
Prepared by

Eva Maria HAGER

06. NOV. 2019

Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR

06. NOV. 2019

Datum / Date: