CERTIFICATE OF ANALYSIS

CHROMOGENIX Instrumentation Laboratory
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2021-05
PRODUCT NUMBER:	82266863	
LOT NUMBER:	N0596176	

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	DISPOSITION
Analysis with COATEST APC R:		MIN	MAX	UNITS		PASS/FAIL
APT Time	$CaCl_2$	27	44	seconds	41	PASS
	APC/$CaCl_2$	--	101	seconds	92	PASS
APC Ratio		--	2.3	--	2.2	PASS
Analysis with COATEST APC R V:						
APT Time	$CaCl_2$	27	40	seconds	37	PASS
APC Ratio		1.4	1.9	--	1.8	PASS

We hereby certify that the above product(s):
· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
· meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel (PRINT) N. M. Patel (SIGNATURE) DATE: 7/15/2019 7-15-19

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz (PRINT) Debra Lee Harz (SIGNATURE) DATE: 7/15/2019 7/15/19

TITLE: Technical Supervisor

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000