

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	CoaTest SP4 FVIII Kit	EXPIRATION DATE:	2021-04
PRODUCT NUMBER:	82409463		
LOT NUMBER:	N0595852		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0087	PASS
Intercept	--	0.250	--	0.140	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.045	PASS
Intercept	--	0.406	--	0.184	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS:

DESCRIPTION	LOT NUMBER	VOLUME
COATEST SP FVIII CACL2	N0595683	6 mL
COATEST SP FVIII FACTOR IXa + FACTOR X	N0595848	3 mL
COATEST SP FVIII S-2765 + I-2581	N0595688	12 mL
COATEST SP FVIII BUFFER	N0495509	20 mL
COATEST SP FVIII PHOSPHOLIPID	N0595689	2 mL

PREPARED BY: Nehal Patel *N.m. Patel* DATE: 6/26/2019 *6-26-19*
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 6/26/2019 *6/26/19*
 PRINT SIGNATURE

TITLE: Technical Supervisor