



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ6M00.01

Att.06

6M93BB0.03

## TECHNOCLOT® PT Owren manual 10x10 mL

REF 5005037  
LOT 6M93BB0.03  
⌚ 2021-07-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Cap color	white	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Kontrollwiederfindung Recovery of controls	<b>Control N</b> LOT 1P71B01	0.92 - 1.16 INR	1.03 INR	OK
	<b>Control A</b> LOT 3P71B01	1.89 - 2.39 INR	2.11 INR	OK

Durchgeführt von  
Prepared by

  
Stefan STEINER

21. NOV. 2019

Datum / Date:

Freigabe von  
Approved by

  
Suzana GRESAK PODAKAR

21. NOV. 2019

Datum / Date: