



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0D00AZ

TECHNOTHROMBIN® TGA CAL Set

REF 5006345

LOT 0D93B00.02

🕒 2020-11-30

Inhalt Content	Reagenz Reagent	LOT	🕒
1x3mL	TGA CAL BUF	9P91B00	2021-05-31
1x0.5mL	TGA CAL	5T81B	2020-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kalibrierung / Calibration gegen / against WHO 01/580		report	742	OK	
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197074	Regression der Bezugskurve / regression of the calibration curve	≥ 0.99	0.99	OK	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z91B00	TGA RCH Reagent LOT 7F92B00	417 – 695	676.9	OK
	TGA Control Low (nM Peak Thrombin) LOT 2891B00	TGA RCH Reagent LOT 7F92B00	248 - 413	385.5	OK

Durchführung von
Prepared by

Michael KAFKA

28. NOV. 2019

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

28. NOV. 2019

Datum / Date: