
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0300.01
---	---	-----------

Technoview® Rivaroxaban CAL Set High 5x1 mL

REF 5090171
LOT 0391B00.01
 2021-06-30

Inhalt Content	Reagenz Reagent	LOT	
1x1 mL	Technoview Rivaroxaban CAL1 High	5181B01	2021-06-30
1x1 mL	Technoview Rivaroxaban CAL2 High	5281B01	2021-06-30
1x1 mL	Technoview Rivaroxaban CAL3 High	5381B01	2021-06-30
1x1 mL	Technoview Rivaroxaban CAL4 High	5481B01	2021-06-30
1x1 mL	Technoview Rivaroxaban CAL5 High	5581B01	2021-06-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	Rivaroxaban Control Medium LOT 3591B00	80.6 – 134.4 ng/mL	96.7	OK
	Rivaroxaban Control High LOT 3381B01	245.2 – 408.6 ng/mL	257.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Stefan STEINER

29. NOV. 2019
Datum / Date:

**Freigabe von
Approved by**


Eva Maria HAGER

29. NOV. 2019
Datum / Date: