
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3400.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control Low 5x1 mL

REF 5090172
LOT 3482B01.02
 2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		47.3 – 87.8 ng/mL	67.5	OK
Verwendbarkeit Usability Ceveron Alpha®	TECHNOCHROM® anti-Xa	47.3 – 87.8 ng/mL	47.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Stefan STEINER

02. AUG. 2019

Datum / Date:

**Freigabe von
Approved by**


Riham EL KAMMAR

02. AUG. 2019

Datum / Date: