

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control High 5x1 mL

REF	5090174
LOT	3381B01.01
	2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		245.2 – 408.6 ng/mL	326.9	
Verwendbarkeit Usability Ceveron® alpha	TECHNOCHROM® anti-Xa	245.2 – 408.6 ng/mL	257.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

2 9. NOV. 2019
Datum / Date:

Freigabe von
Approved by


Eva Maria HAGER

2 9. NOV. 2019
Datum / Date: