



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRA00AZ

TECHNOZYM[®] vWF:Ag ELISA Kit 96T

REF 5450201
LOT RA93B00.01
2021-05-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:Ag Plate	RNF91B0	2021-10-31
1 x 0.3mL	Ag Conjugate	RNH81B1	2021-06-30
1x 80mL	Wash buffer Concentrate	RNA93B0	2021-09-30
1 x 90mL	Incubation buffer	RNB92B0	2021-07-31
1 x 12mL	TMB Substrate	RKN92B0	2021-07-31
1 x 12mL	Stop solution	RKK92B0	2021-06-30
1 x 0.5mL	vWF Calibrator 1	1ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 2	2ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 3	3ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 4	4ND91B0	2021-05-31
1 x 0.5mL	vWF Calibrator 5	5ND92B0	2021-06-30
1 x 0.5mL	vWF High Control	AND91B0	2021-05-31
1 x 0.5mL	vWF Low Control	BND91B0	2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.39	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND91B0)	0.92 – 1.54 IU/mL	1.16	OK
	Low Control (BND91B0)	0.22 – 0.50 IU/mL	0.36	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

25.05.2019

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

25.05.2019

Datum / Date: