

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	FXa, 10 x 71 nkat	EXPIRATION DATE:	2022-01
PRODUCT NUMBER:	82098539		
LOT NUMBER:	N0797522		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	74	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.0	73.0	nkat/vial	69.3	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 8-22-19
PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 8/22/19
PRINT SIGNATURE

TITLE: QC Technical Supervisor