

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE:	2022-09
PRODUCT NUMBER:	82032439		
LOT NUMBER:	N0998410		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 10-2-19
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 10-3-19
 PRINT SIGNATURE

TITLE: QC Technical Supervisor