



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRL00AZ

TECHNOZYM<sup>®</sup> t-PA Antigen EDTA ELISA Kit

REF TC12007

LOT RL94B00.01

🕒 2021-06-30

| Inhalt<br>Content | Reagenz<br>Reagent      | LOT     | 🕒          |
|-------------------|-------------------------|---------|------------|
| 1x                | GN-Anti t-PA Plate      | RGN92B0 | 2021-10-31 |
| 1x0.3mL           | KM-POX anti t-PA        | RKM81B2 | 2021-08-31 |
| 1x80mL            | Wash buffer Concentrate | RNA93B0 | 2021-09-30 |
| 1x90mL            | Incubation buffer       | RNB92B0 | 2021-07-31 |
| 1x12mL            | TMB Substrat 12mL       | RKN92B0 | 2021-07-31 |
| 1x12mL            | Stop solution           | RKK92B0 | 2021-06-30 |
| 1x20mL            | Dilution buffer         | RAD91B0 | 2020-08-31 |
| 1x0.5mL           | Calibrator 1 t-PA       | 1AV82B1 | 2021-08-31 |
| 1x0.5mL           | Calibrator 2 t-PA       | 2AV81B1 | 2021-08-31 |
| 1x0.5mL           | Calibrator 3 t-PA       | 3AV81B1 | 2021-08-31 |
| 1x0.5mL           | Calibrator 4 t-PA       | 4AV81B1 | 2021-08-31 |
| 1x0.5mL           | Calibrator 5 t-PA       | 5AV81B1 | 2021-08-31 |
| 1x0.5mL           | Control High t-PA       | AAV82B1 | 2021-08-31 |
| 1x0.5mL           | Control Low t-PA        | BAV82B1 | 2021-08-31 |

| Analytische Prozedur<br>Analytical Procedure |                        | Anforderung<br>Requirement | Ergebnis<br>Result |    |
|--|------------------------|----------------------------|--------------------|----|
| OD   | Calibrator 1           | > 1.0                      | 1.69               | OK |
|  | Calibrator 5           | < 0.1                      | 0.02               | OK |
| Verwendbarkeit<br>Usability                  | High Control (AAV82B1) | 10.79 – 17.99 ng/mL        | 14.73              | OK |
|  | Low Control (BAV82B1)  | 2.01 – 4.69 ng/mL          | 3.84               | OK |

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

29. OKT. 2019

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

29. OKT. 2019

Datum / Date: