



Certificate of Analysis

Factor Xla Control

REF	Catalog number	1188
LOT	Batch code	HHR1934
	Use by	2024 08
	Upper limit of temperature	-20°C


Activity after reconstitution with 4.0 mL water:

Acceptance range
9.0 – 13.4 mIU/mL

The above Factor Xla activity apply only to Lot#HHR1934

The calibration was performed using the Rox Factor Xla kit (Ref#110050). Factor Xla Calibrator (Ref#1199), potency assigned vs the 1st International standard for Activated Blood Coagulation Factor XI (FXIa), Human, NIBSC code 13/100, was used as calibrator.

This Lot# has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications. All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.



Pia Bryngelhed
Quality Manager, Rossix AB

October 02, 2019

Date



Rossix AB
Taljegardsgatan 3B
SE-431 53 Molndal
Sweden

Phone: +46 31 706 8965
Fax: +46 31 706 8966
info@rossix.com
www.rossix.com