



# Certificate of Analysis

## Factor Xla Calibrator

<b>REF</b>	Catalog number	1199
<b>LOT</b>	Batch code	CAR1337
	Use by	2024 09
	Upper limit of temperature	-20°C


Activity after reconstitution with 4.0 mL water:

**89 mIU/mL**

The above Factor Xla activity apply only to Lot#**CAR1337**.

The calibration has been made against the 1<sup>st</sup> International standard for Activated Blood Coagulation Factor XI (FXIa), Human, NIBSC code 13/100 using the Rox Factor Xla kit Ref 110050.

This Lot# has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications. All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.

  
Pia Bryngelhed  
Quality Manager, Rossix AB

**October 10, 2019**

Date



Rossix AB  
Taljegardsgatan 3B  
SE-431 53 Molndal  
Sweden

Phone: +46 31 706 8965  
Fax: +46 31 706 8966  
[info@rossix.com](mailto:info@rossix.com)  
[www.rossix.com](http://www.rossix.com)