

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ8F00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA SUB 5x1.5mL

REF 5006235

LOT 8F91B00.03

 2021-04-30

Inhalt Content	Reagenz Reagent	LOT	
5x1.5 mL	TGA Substrat	8F91B00	2021-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z91B00	TGA RC High LOT 7F92B00	417 - 695	556	OK
	TGA Control Low (nM Peak Thrombin) LOT 2891B00	TGA RC High LOT 7F92B00	248 - 413	358	OK

Durchführung von  
Prepared by

Michael KAFKA 

15. JAN. 2020

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI 

16. JAN. 2020

Datum / Date: