



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ0U00.02

## AK-Calibrant 4 x 1 mL

REF 5010004  
LOT 0U83B00.06  
🕒 2021-09-30

Inhalt Content	Reagenz Reagent	LOT	🕒
1x1mL	AK Calibrant A	1N84B00	2021-09-30
1x1mL	AK Calibrant B	3L88B00	2021-09-30
1x1mL	AK Calibrant C	3M8DB00	2021-10-31
1x1mL	AK Calibrant D	3N8EB00	2021-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P71B01 Technoplastin HIS	96.7 – 121.9 % 0.80 – 1.00 INR	104.8 0.97	OK OK
	CCA LOT 3P71B01 Technoplastin HIS	24.2 – 30.6 % 2.68 – 3.38 INR	28.2 2.94	OK OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
Stefan STEINER

17. JAN. 2020

Datum / Date:

Freigabe von  
Approved by

  
Eva Maria HAGER

17. JAN. 2020

Datum / Date: