

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3F00.01
---	---	-----------

TECHNOVIEW[®] LMW Low CONT 5x1 mL

REF	5090042
LOT	3F91B00.01
	2021-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.17 – 0.31 IU/mL	0.24	OK
Verwendbarkeit Usability Ceveron [®] Alpha	Anti-Xa LMWH	0.17 – 0.31 IU/mL	0.30	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Stefan STEINER

07. JAN. 2020

Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR

07. JAN. 2020

Datum / Date: