



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3H00.02

TECHNOVIEW® LMW CONT H 5x1 ml

REF 5090046  
LOT 3H82B00.02  
⌚ 2020-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Richtwert Expected value Ceveron® alpha	Anti-Xa	0.77 – 1.28 IU/mL		
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa	0.77 – 1.28 IU/mL	1.05	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Eva Maria HAGER

11. NOV. 2019  
Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

11. NOV. 2019  
Datum / Date: