
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3Q00.01
---	---	-----------

**TECHNOVIEW® Organan CONT Low 5x1 mL**

REF	5090112
LOT	3Q81B00.02
	2020-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.43 – 0.79 IU/mL	0.61	OK
Verwendbarkeit Usability Ceveron® alpha	Technochrom® anti Xa Organan	0.43 – 0.79 IU/mL	0.56	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
**Stefan STEINER**

12. SEP. 2019  
Datum / Date:

**Freigabe von**  
**Approved by**

  
**Riham EL KAMMAR**

12. SEP. 2019  
Datum / Date: