



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

3151B01.01

Technoview® Argatroban CONT Low 5x1mL

REF	5090142
LOT	3151B01.01
	2020-08-31

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Cap color	red confirmed OK
	Vial type	clear glass vial confirmed OK
	Vial label	according to packaging SOP confirmed OK
	Vial barcode	according to packaging SOP confirmed OK
	Box label	according to packaging SOP confirmed OK
	Box barcode	according to packaging SOP confirmed OK
	Package insert	present confirmed OK
	Batch table	present confirmed OK
	Appearance before reconstitution	buff colored plug of lyophilized material confirmed OK
	Appearance after reconstitution	buff colored liquid confirmed OK
Virologie / Virology The testing methods applied were FDA-approved or CE marked.	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab Negative Negative Negative	Negative Negative Negative OK
Kontrollwiederfindung am Ceveron® Alpha Recovery of controls on Ceveron® Alpha	TECHNOVIEW® Argatroban LOT 0X83B00 0.48 - 0.90 µg/mL	0.65 µg/mL OK

Durchgeführt von
Prepared by

Suzana GRESAK PODAKAR
Suzana Gresak Podakar

12. JULI 2019

Datum / Date:

Freigabe von
Approved by

Stefan STEINER
Stefan Steiner

17. JULI 2019

Datum / Date: