



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0500.01

Technochrom AT III manual method Kit ~39 T.

REF 5340225

LOT 0592B00.01

2021-05-31

Inhalt Content	Reagenz Reagent	LOT	
1x	Substrate Th-1 10µmol QC	8B82B00	2021-07-31
1x	Technochrom ATIII Reagent A1 20µmol	7U61B02	2021-05-30
1x	Technochrom ATIII Puffer A1 100mL	9871B00	2022-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	CCN LOT 1P71B01	75.8 – 126.3%	94.8	OK
	CCA LOT 3P71B01	33.3 – 55.4%	49.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Michael KAFKA

18. JULI 2019

Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR

18. JULI 2019

Datum / Date: