

**CERTIFICATE OF ANALYSIS**



PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2021-07
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0797203	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.57	--	0.41	PASS
Intercept	---	0.12	--	0.06	PASS
r	0.990	---	--	0.997	PASS
LOW RANGE					
Slope	2.8	6.9	--	4.4	PASS
Intercept	---	0.21	--	0.09	PASS
r	0.990	---	--	0.997	PASS

**KIT COMPONENTS**

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0797201	3 mL
COAMATIC FVIII S-2785 + I-2581	N0797055	6 mL
COAMATIC FVIII BUFFER	N0797052	24 mL

· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Maria Oettinger *Maria Oettinger*  
 PRINT SIGNATURE

DATE: 8-13-19

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*  
 PRINT SIGNATURE

DATE: 8-13-19

TITLE: QC Technical Supervisor