

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
 A Werfen Company

PRODUCT NAME:	S-2302, 25 mg Substrate	EXPIRATION DATE:	2022-12
PRODUCT NUMBER:	82034039		
LOT NUMBER:	N0898087		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 9-13-19
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 9/13/19
 PRINT SIGNATURE

TITLE: QC Technical Supervisor