



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRE00AZ

TECHNOZYM® PAI-1 Actibind ELISA Kit 96T

REF TC16075

LOT RE92B00.01

🕒 2021-02-28

Inhalt/ Content	Reagenz/Reagent	LOT	🕒
1 x	anti t-PA coated plate Actibind	RGP92B0	2021-07-31
1 x 0.3 mL	anti PAI-1 Conjugate	RKQ81B1	2021-02-28
1 x 80 mL	Wash buffer Concentrate	RNA92B0	2021-06-30
1 x 90 mL	Incubation buffer	RNB92B0	2021-07-31
1 x 12 mL	TMB Substrat	RKN92B0	2021-07-31
1 x 12 mL	Stop solution	RKK92B0	2021-06-30
1 x 0.2 mL	Calibrator 1 active PAI-1	1DW95B0	2021-09-30
1 x 0.2 mL	Calibrator 2 active PAI-1	2DW92B0	2021-07-31
1 x 0.2 mL	Calibrator 3 active PAI-1	3DW92B0	2021-07-31
1 x 0.2 mL	Calibrator 4 active PAI-1	4DW92B0	2021-07-31
1 x 0.2 mL	Calibrator 5 active PAI-1	5DW91B0	2021-07-31
1 x 0.2 mL	High Control active PAI-1	ADW91B0	2021-07-31
1 x 0.2 mL	Low Control active PAI-1	BDW91B0	2021-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	1 – 2.5	1.42	OK
	OD Cal 5	< 0.2	0.03	OK
	High Control (ADW91B0)	12.80 – 21.34 IU/mL	19.48	OK
	Low Control (BDW91B0)	2.49 – 5.81 IU/mL	2.78	OK

Durchführung von
Prepared by

Michael KAFKA

25. SEP. 2019

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

25. SEP. 2019

Datum / Date: