



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3C00.02

TECHNOVIEW UFH Control L 5x1 mL

REF 5090072

LOT 3C81B01.02

 2021-06-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® Alpha	Anti-Xa UFH	0.11 – 0.20 IU/mL	0.12	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

05. FEB. 2020
Datum / Date:

Freigabe von
Approved by


Christoph ENNE

05. FEB. 2020
Datum / Date: