

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3D00.01
-----------------------------------------------------------------------------------	-------------------------------------------------------	-----------

TECHNOVIEW® Apixaban CONT High 5x1 mL

REF 5090270
LOT 3D81B01.01
 2021-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		229.3 – 382.1 ng/mL	305.7	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	229.3 – 382.1 ng/mL	286.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

14. FEB. 2020
 Datum / Date:

Freigabe von
Approved by


Eva Maria HAGER

14. FEB. 2020
 Datum / Date: