

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance V	EXPIRATION DATE:	2021-07
PRODUCT NUMBER:	82312063		
LOT NUMBER:	N0101075		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	Pass/Fail
	MIN	MAX			
<b>APT-time on ACL</b>					
Diluted Control Plasma Level 1	28	47	seconds	32	PASS
Diluted Control Plasma Level 2	30	50	seconds	34	PASS

<b>APC-V ratio on ACL</b>					
Control Plasma Level 1	2.8	3.2	--	3.1	PASS
Control Plasma Level 2	1.4	1.9	--	1.8	PASS

**KIT COMPONENTS:**

NAME	LOT NUMBER	VOLUME
COATEST APC Resistance CaCl2	N1290520	8 mL
COATEST APC Resistance APTT reagent	N0100837	16 mL
COATEST APC Resistance APC/CaCl2	N0100859	2 mL
COATEST APC Resistance Control Plasma Level 1	N0100996	1 mL
COATEST APC Resistance Control Plasma Level 2	N0898224	1 mL
COATEST APC Resistance V-DEF Plasma	N1098981	4 mL

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel *N.M. Patel* DATE: 1-29-20  
PRINT SIGNATURE DATE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 1/29/2020  
PRINT SIGNATURE DATE

TITLE: Technical Supervisor