



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ8F00AZ

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230

LOT 8F01C00.01

⌚ 2022-02-28

Inhalt Content	Reagenz Reagent	LOT	⌚
50x1.5mL	TGA Substrat	8F01C00	2022-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z01C00	TGA RC Low LOT 7D01C00 253 - 421	274.1	OK
		TGA RC High LOT 7F92B00 435 - 725	582.6	OK
	TGA Control Low (nM Peak Thrombin) LOT 2801C00	TGA RC Low LOT 7D01C00 164 - 274	190.1	OK
		TGA RC High LOT 7F92B00 225 - 375	300.1	OK

Durchführung von
Prepared by

Rikret NASUFI *Fahed Nasufi*

17. MRZ. 2020

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

17. MRZ. 2020

Datum / Date: