

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	
		DQ8F00AZ

TECHNOTHROMBIN[®] TGA SUB 5x1.5 mL

REF 5006235

LOT 8F01C00.01

 2022-02-28

Inhalt Content	Reagenz Reagent	LOT	
5x1.5 mL	TGA Substrat	8F01C00	2022-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z01C00	TGA RC Low LOT 7D01C00	253 – 421	310	OK
	TGA Control Low (nM Peak Thrombin) LOT 2801C00	TGA RC Low LOT 7D01C00	164 – 274	218	OK

Durchführung von
Prepared by

Michael KAFKA

16. MRZ. 2020

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI *Fikret Nasufi*

16. MRZ. 2020

Datum / Date: