

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2800AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Control Low 5x1mL

REF 5006330
LOT 2801C00.01
 2021-01-31

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control Low	2801C00	2021-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls TGA RC Low (nM Peak Thrombin) LOT 7D01C00	164 - 274	218	OK

Durchführung von
Prepared by

Michael KAFKA 

16. MRZ. 2020
Datum / Date:

Freigabe von
Approved by

Ikret NASUFI 

16. MRZ. 2020
Datum / Date: