


	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	 
		DQ1P00.01

### Coagulation Control N 5x1 mL

REF      5020040  
LOT      1P02C00.01  
      2021-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	Siron LS	22.2 – 27.4 sec	26.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

Michael KAFKA



17. MRZ. 2020

Datum / Date:

**Freigabe von**  
**Approved by**

Alessandro BENIGNI



17-Mrz-2020

Datum / Date: