

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control High 5x1 mL

REF 5090174
LOT 3391B00.01
 2021-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		235.9 – 393.1 ng/mL	314.5	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW® anti-Xa	235.9 – 393.1 ng/mL	322.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Stefan
Stefan STEINER

13. MRZ. 2020

Datum / Date:

Freigabe von
Approved by

Eva Maria Hager
Eva Maria HAGER

13. MRZ. 2020

Datum / Date: