



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3900.01

TECHNOVIEW® Apixaban CONT Low 5x1 ml

REF 5090271

LOT 3981B01.01

🕒 2021-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		64.2 – 119.2 ng/mL	91.7	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	64.2 – 119.2 ng/mL	76.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

14. FEB. 2020
Datum / Date:

Freigabe von
Approved by


Eva Maria HAGER

14. FEB. 2020
Datum / Date: