



technoclone

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1R00.01

Coagulation Reference, 5x1mL

REF 5220110

LOT 1R91B00.02

 2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	F IX Siron CCN LOT 1P71B01	95.2 – 158.7 %	110.9	OK
	F IX Siron CCA LOT 3P91B00	12.0 – 20.0 %	15.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

19. MRZ. 2020

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI *Fikret Nasufi*

19. MRZ. 2020

Datum / Date: